



Plasmid DNA concentration and desalination

User manual

NucleoSnap Finisher Midi
NucleoSnap Finisher Maxi

January 2016 / Rev. 01

Plasmid DNA concentration and desalination

Protocol-at-a-glance (Rev. 01)






	Midi		Maxi
	NucleoSnap Finisher Midi		NucleoSnap Finisher Maxi
1 Adjust binding conditions	2.5 mL Buffer FB Mix		7.5 mL Buffer FB Mix
2 Filtrate precipitate	Connect NucleoSnap Column to vacuum Load sample -0.3 bar*		Connect NucleoSnap Column to vacuum Load sample -0.3 bar*
3 Wash precipitate	2 mL Buffer A4 -0.3 bar*		4 mL Buffer A4 -0.3 bar*
4 Separate column	Remove upper part and discard		Remove upper part and discard
5 Dry silica membrane	11,000 x g, 1 min		11,000 x g, 1 min
6 Elute DNA	200 µL H ₂ O-EF 11,000 x g, 1 min		500 µL H ₂ O-EF 11,000 x g, 1 min

Table of contents

1	Components	4
1.1	Kit contents	4
1.2	Reagents and equipment to be supplied by user	4
2	Kit specifications	5
2.1	About NucleoSnap Finisher kits	5
2.2	Setup of NucleoSnap Finisher Columns	5
2.3	Elution procedures	6
3	Storage conditions and preparation of working solutions	7
4	Safety instructions	7
5	Protocol for plasmid concentration and desalination	8
5.1	Plasmid purification from anion exchange prep eluates	8
5.2	Support Protocol for purification of DNA from samples with low ionic strength	10
6	Appendix	11
6.1	Troubleshooting	11
6.2	Ordering information	12
6.3	Product use restriction / warranty	12

1 Components

1.1 Kit contents

REF	NucleoSnap Finisher Midi		NucleoSnap Finisher Maxi	
	10 preps 740434.10	50 preps 740434.50	10 preps 740435.10	50 preps 740435.50
Buffer FB	30 mL	150 mL	90 mL	400 mL
Buffer A4 (Concentrate)	6 mL	25 mL	2 x 6 mL	2 x 25 mL
H ₂ O-EF	13 mL	30 mL	30 mL	60 mL
NucleoSnap Finisher Columns	10	50	10	50
Collection Tubes	10	50	10	50
User manual	1	1	1	1

1.2 Reagents and equipment to be supplied by user

Reagents

- 96–100 % ethanol

Equipment

- Vacuum manifold with Luer adapters (NucleoVac Mini Adapters)
- Vacuum pump capable of reaching -0.3 bar (9–10 in. Hg)
- Centrifugation tubes (2 mL)
- Centrifuge capable of reaching $\geq 10,000 \times g$
- Pipettes and pipette tips for 0.1–1 mL and 0.5–10 mL

2 Kit specifications

- The **NucleoSnap Finisher** kits are specifically designed for the fast and parallel purification and concentration of up to 2 mg of plasmid DNA from anion exchange preparation eluates.
- The **NucleoSnap Finisher Midi** kit can be used in combination with the NucleoBond® Xtra Midi and the NucleoBond® PC 100 kits.
- The **NucleoSnap Finisher Maxi** kit can be used in combination with the NucleoBond® Xtra Maxi and the NucleoBond® PC 500 kits.
- NucleoBond® PC 2000 eluates require additional Buffer FB (see ordering information, section 6.2) and can be used according to the NucleoSnap Finisher Maxi protocol. Filtration times will increase when more than 1.5 mg DNA are being purified.
- Purification of DNA from samples with low ionic strength like pure or slightly buffered water is possible according to protocol 5.2 but will require additional Buffer PL3 (see ordering information, section 6.2).

2.1 About NucleoSnap Finisher kits

NucleoSnap Finisher kits supersede the commonly used and time consuming isopropanol precipitation and centrifugation steps by allowing the rapid DNA precipitation and concentration of large volumes under vacuum.

DNA is precipitated by Precipitation Buffer FB and filtered by the specially designed matrix of the novel **NucleoSnap Finisher Columns**, followed by a washing step with ethanolic Wash Buffer A4 to remove salts and impurities.

To enable the use of a table top centrifuge, the **NucleoSnap Finisher Column** is composed of an upper funnel component which can be removed from the lower Mini spin column part by a simple break action. Then the Mini spin column can be centrifuged in a standard table top centrifuge to remove residual ethanol.

Afterwards, plasmid DNA can be eluted in supplied endotoxin-free H₂O-EF. The applied elution volume can be adjusted from 100 µL to 1.5 mL according to the expected amount of plasmid DNA to ensure optimal yield and concentration. See chapter 2.3 for recommendations concerning the optimal elution volume.

2.2 Setup of NucleoSnap Finisher Columns

NucleoSnap Finisher Columns are connected to a vacuum manifold either by a direct Luer connection or by the use of a NucleoVac Mini Adapter or NucleoVac Stop-cock (see ordering information, section 6.2). A stop-cock is useful to switch off vacuum selectively when using a large number of columns at the same time to prevent a pressure loss through empty columns and excessive filtration of potentially contaminated air.

The space between each two used inlets of the vacuum manifold should be sufficient not to bend or dislodge **NucleoSnap Finisher Columns** attached to the vacuum manifold.

The **NucleoSnap Finisher Columns** consist of one piece but can be divided into two parts: a lower Mini spin column part and an upper funnel part. Handle the columns carefully to prevent accidental damage to the predetermined breaking point!



Buffer FB is viscous. Use of **reverse pipetting** is recommended to ensure accurate volumes. Reverse pipetting is done by pressing down the pipette's plunger button all the way down to the second stop before slowly aspirating the Buffer FB until the plunger button rests again in the starting position. The buffer volume inside the pipette tip is larger than set now, so when dispensing the Buffer FB to the anion exchange eluates be sure to dispense to the first stop only! Liquid remaining in the pipette tip can be dispensed back to the buffer container.

For further details concerning the reverse pipetting technique and liquid handling of viscous fluids you may also check your pipette manufacturer's information material.

2.3 Elution procedures

Total yield depends on the final DNA concentration in the eluates.

A higher concentration than 2 µg/µL of DNA is difficult to elute from a silica spin column. As a result, DNA will not elute completely if the elution buffer is already saturated with plasmid DNA.

To prevent reduced total yield as a side-effect of high concentration it is recommended to measure the DNA content of the anion exchange eluate and to choose the total elution volume accordingly to gain a final concentration of 1–2 µg/µL.

- To increase **concentration**, perform 2 elution steps by reloading the eluate (generally recommended).
- To increase **yield**, perform 2 x 500 µL elution steps with fresh nuclease-free H₂O-EF for a total elution volume of 1 mL. Measure DNA content in both fractions separately.

In general it is advantageous to incubate the nuclease-free H₂O-EF on the membrane at room temperature or elevated temperatures (e.g., 50–70 °C) for 1–5 min.

Recommended elution volumes according to expected yield

Kit	DNA yield	Recommended volume for elution
NucleoBond® PC 100	up to 100 µg	100–150 µL
NucleoBond® Xtra Midi	up to 500 µg	200–500 µL
NucleoBond® PC 500	up to 500 µg	200–500 µL
NucleoBond® Xtra Maxi	up to 1500 µg	2 x 500 µL
NucleoBond® PC 2000	up to 2000 µg	2 x 500 µL

3 Storage conditions and preparation of working solutions

All kit components can be stored at room temperature (18–25 °C) and are stable for at least two years.

Before starting any **NucleoSnap Finisher** protocol prepare the following:

- **Wash Buffer A4:** Add the given volume of ethanol (96–100 %) as indicated on the bottle or in the table below to **Buffer A4 (Concentrate)** before first use. Mark the label on the bottle to indicate that the ethanol is added. Prepared Buffer A4 is stable at room temperature (18–25 °C) for at least one year.

REF	NucleoSnap Finisher Midi		NucleoSnap Finisher Maxi	
	10 preps 740434.10	50 preps 740434.50	10 preps 740435.10	50 preps 740435.50
Buffer A4 (Concentrate)	6 mL Add 24 mL ethanol	25 mL Add 100 mL ethanol	2 x 6 mL Add 24 mL ethanol to each bottle	2 x 25 mL Add 100 mL ethanol to each bottle

4 Safety instructions

The **NucleoSnap Finisher** kits do not contain hazardous components.

Wear gloves and goggles and follow the manufacturers' safety instructions for using the vacuum manifold!

5 Protocol for plasmid concentration and desalination

5.1 Plasmid purification from anion exchange prep eluates

Before starting the preparation:

- Check if Buffer A4 was prepared according to section 3.

All vacuum steps are performed with a pressure of about -0.3 bar* (10 in. Hg).

Midi**Maxi**

1 Adjust precipitation conditions

Add **0.5 volumes** of **Buffer FB** to anion exchange eluate.

Reverse pipetting is recommended (see chapter 2.2).

2.5 mL**7.5 mL**

Vortex for 5 s.

2 Filtrate precipitate

Connect the **NucleoSnap Finisher Column** to a vacuum manifold and load mixture from step 1 onto the column.

Apply vacuum (-0.3 bar*) until the solution has completely passed and then turn vacuum off.

3 Wash precipitate

Add **Buffer A4** onto the column and apply vacuum (-0.3 bar*) until the solution has completely passed and then turn vacuum off.

2 mL**4 mL**

4 Separate column

Remove the **NucleoSnap Finisher Column** from the vacuum manifold and place the bottom Mini spin column part into a 2 mL Collection Tube (supplied).

Snap off the funnel part from the Mini spin column part (placed in the Collection Tube). Discard the funnel.

5 Dry silica membrane

Centrifuge Mini spin column and Collection Tube for **1 min** at **> 11,000 x g** to remove any residual ethanol.

Discard the Collection Tube and place the Mini spin column in a new 2 mL Eutition tube (not supplied).

*Reduction of atmospheric pressure.

Midi

Maxi

6 Elute DNA

Add **200–500 µL** of nuclease- and endotoxin-free **H₂O-EF** onto the membrane (for recommended elution volumes, see section 2.3) and incubate at room temperature for 1 min.

Centrifuge for **1 min** at **> 10,000 x g**.

Optional: Repeat elution with the eluate as elution buffer for optimal recovery. See chapter 2.3 for further hints.

Recommended for Maxi preps: Place Mini spin column in a new 2 mL elution tube (not supplied). Repeat elution with 500 µL of fresh H₂O-EF as elution buffer for optimal recovery. Check DNA concentration in both fractions.

5.2 Support Protocol for purification of DNA from samples with low ionic strength

Purification of DNA from samples with low ionic strength like DNA in water or slightly buffered solutions requires an adaption of buffer conditions, ionic strength, and pH value. Add 1/3 sample volume of Buffer PL3 (has to be ordered separately, see ordering information, section 6.2) to the sample prior to the addition of Buffer FB.

1 Adjust precipitation conditions

Add **0.33 volumes** of **Buffer PL3** to low ionic strength sample.

E.g., add 3 mL Buffer PL3 to 9 mL sample.

2 Precipitate DNA

Add **0.5 total volumes** of **Buffer FB** to the mixture.

E.g., add 6 mL Buffer FB to 12 mL mixture from step 1 (consisting of 9 mL sample and 3 mL Buffer PL3).

Reverse pipetting is recommended (see chapter 2.2).

Vortex for **5 s**.

Continue with step 2 of protocol 5.1

6 Appendix

6.1 Troubleshooting

Problem	Possible cause and suggestions
No or low DNA yield	<i>No plasmid DNA present in anion exchange eluates.</i>
	<ul style="list-style-type: none"> • Measure DNA yield after anion exchange prep.
	<i>Insufficient amount of Buffer FB added.</i>
	<ul style="list-style-type: none"> • Precipitation Buffer FB is viscous, make sure to add the correct volume.
	<ul style="list-style-type: none"> • Use “reverse pipetting” to avoid inaccurate pipetting of precipitation buffer (see section 2.2)
	<ul style="list-style-type: none"> • Precipitation works best when 0.5 vol of Buffer FB are added to each vol of anion-exchange eluate. When using other volumes than those of the standard procedure, adjust volume of Buffer FB accordingly.
Slow flow rates	<i>Excess plasmid input</i>
	<ul style="list-style-type: none"> • Plasmid DNA is filtrated on top of the filter membrane. Increasing amounts of plasmid DNA will lead to reduced flow rates when more than 1.5 mg DNA have been loaded. Loading DNA amounts higher than 2.5 mg might lead to column clogging.
	<i>Insufficient vacuum force</i>
	<ul style="list-style-type: none"> • The lower the vacuum force the slower the flow rate will be. Use vacuum pumps only that enable a minimum of -0.3 bar.

6.2 Ordering information

Product	REF	Pack of
NucleoSnap Finisher Midi	740434.10	10 preps
	740434.50	50 preps
NucleoSnap Finisher Maxi	740435.10	10 preps
	740435.50	50 preps
Vacuum manifold	740299	1
NucleoVac Stop-Cocks	740298.24	24
NucleoVac Mini Adapter	740297.100	100
Buffer FB	740438.1000	1000 mL
Wash Buffer A4 (Concentrate)	740914.1	200 mL
H ₂ O-EF	740798.1	1000 mL
Buffer PL3	740352.50	50 mL
	740352.1000	1000 mL

6.3 Product use restriction/warranty

NucleoSnap Finisher kit components are intended, developed, designed, and sold FOR RESEARCH PURPOSES ONLY, except, however, any other function of the product being expressly described in original MACHEREY-NAGEL product leaflets.

MACHEREY-NAGEL products are intended for GENERAL LABORATORY USE ONLY! MACHEREY-NAGEL products are suited for QUALIFIED PERSONNEL ONLY! MACHEREY-NAGEL products shall in any event only be used wearing adequate PROTECTIVE CLOTHING. For detailed information please refer to the respective Material Safety Data Sheet of the product! MACHEREY-NAGEL products shall exclusively be used in an ADEQUATE TEST ENVIRONMENT. MACHEREY-NAGEL does not assume any responsibility for damages due to improper application of our products in other fields of application. Application on the human body is STRICTLY FORBIDDEN. The respective user is liable for any and all damages resulting from such application.

DNA/RNA/PROTEIN purification products of MACHEREY-NAGEL are suitable for IN-VITRO-USES ONLY!

ONLY MACHEREY-NAGEL products specially labeled as IVD are also suitable for IN-VITRO-diagnostic use. Please pay attention to the package of the product. IN-VITRO-diagnostic products are expressly marked as IVD on the packaging.

IF THERE IS NO IVD SIGN, THE PRODUCT SHALL NOT BE SUITABLE FOR IN-VITRO-DIAGNOSTIC USE!

ALL OTHER PRODUCTS NOT LABELED AS IVD ARE NOT SUITED FOR ANY CLINICAL USE (INCLUDING, BUT NOT LIMITED TO DIAGNOSTIC, THERAPEUTIC AND/OR PROGNOSTIC USE).

No claim or representations is intended for its use to identify any specific organism or for clinical use (included, but not limited to diagnostic, prognostic, therapeutic, or blood banking). It is rather in the responsibility of the user or - in any case of resale of the products - in the responsibility of the reseller to inspect and assure the use of the DNA/RNA/protein purification products of MACHEREY-NAGEL for a well-defined and specific application.

MACHEREY-NAGEL shall only be responsible for the product specifications and the performance range of MN products according to the specifications of in-house quality control, product documentation and marketing material.

This MACHEREY-NAGEL product is shipped with documentation stating specifications and other technical information. MACHEREY-NAGEL warrants to meet the stated specifications. MACHEREY-NAGEL's sole obligation and the customer's sole remedy is limited to replacement of products free of charge in the event products fail to perform as warranted. Supplementary reference is made to the general business terms and conditions of MACHEREY-NAGEL, which are printed on the price list. Please contact us if you wish to get an extra copy.

There is no warranty for and MACHEREY-NAGEL is not liable for damages or defects arising in shipping and handling (transport insurance for customers excluded), or out of accident or improper or abnormal use of this product; defects in products or components not manufactured by MACHEREY-NAGEL, or damages resulting from such non-MACHEREY-NAGEL components or products.

MACHEREY-NAGEL makes no other warranty of any kind whatsoever, and SPECIFICALLY DISCLAIMS AND EXCLUDES ALL OTHER WARRANTIES OF ANY KIND OR NATURE WHATSOEVER, DIRECTLY OR INDIRECTLY, EXPRESS OR IMPLIED, INCLUDING, WITHOUT LIMITATION, AS TO THE SUITABILITY, REPRODUCTIVITY, DURABILITY, FITNESS FOR A PARTICULAR PURPOSE OR USE, MERCHANTABILITY, CONDITION, OR ANY OTHER MATTER WITH RESPECT TO MACHEREY-NAGEL PRODUCTS.

In no event shall MACHEREY-NAGEL be liable for claims for any other damages, whether direct, indirect, incidental, compensatory, foreseeable, consequential, or special (including but not limited to loss of use, revenue or profit), whether based upon warranty, contract, tort (including negligence) or strict liability arising in connection with the sale or the failure of MACHEREY-NAGEL products to perform in accordance with the stated specifications. This warranty is exclusive and MACHEREY-NAGEL makes no other warranty expressed or implied.

The warranty provided herein and the data, specifications and descriptions of this MACHEREY-NAGEL product appearing in MACHEREY-NAGEL published catalogues and product literature are MACHEREY-NAGEL's sole representations concerning the product and warranty. No other statements or representations, written or oral, by MACHEREY-NAGEL's employees, agent or representatives, except written statements signed by a duly authorized officer of MACHEREY-NAGEL are authorized; they should

not be relied upon by the customer and are not a part of the contract of sale or of this warranty.

Product claims are subject to change. Therefore please contact our Technical Service Team for the most up-to-date information on MACHEREY-NAGEL products. You may also contact your local distributor for general scientific information. Applications mentioned in MACHEREY-NAGEL literature are provided for informational purposes only. MACHEREY-NAGEL does not warrant that all applications have been tested in MACHEREY-NAGEL laboratories using MACHEREY-NAGEL products. MACHEREY-NAGEL does not warrant the correctness of any of those applications.


Last updated: 07/2010, Rev. 03

Please contact:
MACHEREY-NAGEL GmbH & Co. KG
Tel.: +49 (0) 24 21 969 270
e-mail: tech-bio@mn-net.com

Trademarks:

NucleoBond is a registered trademark of MACHEREY-NAGEL GmbH & Co KG

All used names and denotations can be brands, trademarks, or registered labels of their respective owner – also if they are not special denotation. To mention products and brands is only a kind of information (i.e., it does not offend against trademarks and brands and can not be seen as a kind of recommendation or assessment). Regarding these products or services we can not grant any guarantees regarding selection, efficiency, or operation.



www.mn-net.com

MACHEREY-NAGEL



MACHEREY-NAGEL GmbH & Co. KG · Neumann-Neander-Str. 6-8 · 52355 Düren · Germany

DE / International:

Tel.: +49 24 21 969-0
Fax: +49 24 21 969-199
E-mail: info@mn-net.com

CH:

Tel.: +41 62 388 55 00
Fax: +41 62 388 55 05
E-mail: sales-ch@mn-net.com

FR:

Tel.: +33 388 68 22 68
Fax: +33 388 51 76 88
E-mail: sales-fr@mn-net.com

US:

Tel.: +1 484 821 0984
Fax: +1 484 821 1272
E-mail: sales-us@mn-net.com

